

CONFIDENTIAL

10 April 62
ent

ENG 142-

Chief, Operations Division, Medical Staff

Chief, Engineering Staff, CC

Status of Radiological Survey and Radiation Detection Instruments
[redacted] Quarterly Report

25X1

1. During January, February and March 1962, the radiological survey and radiation detection instruments [redacted] were tested to determine their operational status. A total of twelve instruments were found to be inoperative. All instruments, except five were repaired and placed in operating condition.

25X1

2. The five instruments mentioned above are inoperative because of dead batteries and defective power cables. These instruments will be placed in operation as soon as batteries and cables are received.

3. A partial shipment of batteries and components has been received. Battery replacement and calibration of the instruments will begin as soon as the battery shipments are complete.

4. A total of sixteen man-hours has been expended in the last quarter year.

[redacted]

25X1

R&D/Lab/JWM/rkb (10 April 1962)

Distribution:

- Original and 1 - Addressee
1 - Lab Subject
1 - OC-E Chrono
1 - Dev/s
1 - Lab Pending

DOC	5	5-12-80	008632
ENG COMP			
ORIG CLASS	C	1	c
CUST	22	2010	

CONFIDENTIAL